

University of Dundee

## Reduction of bacterial adhesion on titanium-doped diamond-like carbon coatings

Zhao, Ying-Yu; Zhao, Buyun; Su, Xueju ; Zhang, Shuai; Wang, Su; Keatch, Robert

*Published in:*  
Biofouling

*DOI:*  
[10.1080/08927014.2017.1403592](https://doi.org/10.1080/08927014.2017.1403592)

*Publication date:*  
2018

*Licence:*  
CC BY

*Document Version*  
Publisher's PDF, also known as Version of record

[Link to publication in Discovery Research Portal](#)

### *Citation for published version (APA):*

Zhao, Y-Y., Zhao, B., Su, X., Zhang, S., Wang, S., Keatch, R., & Zhao, Q. (2018). Reduction of bacterial adhesion on titanium-doped diamond-like carbon coatings. *Biofouling*, 34(1), 26-33.  
<https://doi.org/10.1080/08927014.2017.1403592>


### General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

### Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.



## Reduction of bacterial adhesion on titanium-doped diamond-like carbon coatings

Ying-Yu Zhao, Buyun Zhao, Xueju Su, Shuai Zhang, Su Wang, Robert Keatch & Qi Zhao

To cite this article: Ying-Yu Zhao, Buyun Zhao, Xueju Su, Shuai Zhang, Su Wang, Robert Keatch & Qi Zhao (2018) Reduction of bacterial adhesion on titanium-doped diamond-like carbon coatings, *Biofouling*, 34:1, 26-33, DOI: [10.1080/08927014.2017.1403592](https://doi.org/10.1080/08927014.2017.1403592)

To link to this article: <https://doi.org/10.1080/08927014.2017.1403592>



© 2018 The Author(s). Published by Informa UK Limited, trading as Taylor & Francis Group



Published online: 15 Jan 2018.



Submit your article to this journal [↗](#)



View related articles [↗](#)



View Crossmark data [↗](#)



## Reduction of bacterial adhesion on titanium-doped diamond-like carbon coatings

Ying-Yu Zhao<sup>a</sup>, Buyun Zhao<sup>b</sup>, Xueju Su<sup>c,d</sup>, Shuai Zhang<sup>d</sup>, Su Wang<sup>d</sup>, Robert Keatch<sup>d</sup> and Qi Zhao<sup>d</sup>

<sup>a</sup>Rehabilitation Medicine Department, the First Hospital of China Medical University, Shenyang, PR China; <sup>b</sup>Medical Research Council Laboratory of Molecular Biology, University of Cambridge, Cambridge, UK; <sup>c</sup>College of Zhuoyue, Bohai University, Jinzhou, PR China; <sup>d</sup>Department of Mechanical Engineering, University of Dundee, Dundee, UK

### ABSTRACT

A range of titanium doped diamond-like carbon (Ti-DLC) coatings with different Ti contents were prepared on stainless steel substrates using a plasma-enhanced chemical vapour deposition technique. It was found that both the electron donor surface energy and the surface roughness of the Ti-DLC coatings increased with increasing Ti contents in the coatings. Bacterial adhesion to the coatings was evaluated against *Escherichia coli* WT F1693 and *Pseudomonas aeruginosa* ATCC 33347. The experimental data showed that bacterial adhesion decreased with the increases of the Ti content, the electron donor surface energy and surface roughness of the coatings, while the bacterial removal percentage increased with the increases of these parameters. The Ti-DLC coatings reduced bacterial attachment by up to 75% and increased bacterial detachment from 15 to 45%, compared with stainless steel control.

### ARTICLE HISTORY

Received 22 June 2017  
Accepted 5 November 2017

### KEYWORDS

Bacterial adhesion; DLC coatings; surface energy; contact angle; biomaterials

## Introduction

The high incidence of infections caused by the use of metallic biomedical implants and devices, including bone fracture fixation pins and hip and knee replacement, has a severe impact on human health and health care costs (Bociaga et al. 2015; Percival et al. 2015; Ceroni et al. 2016; Dunne et al. 2017). Prevention of bacterial adhesion on the surfaces of medical devices will have a major impact in preventing medical device-related infections. An effective and desired approach reducing bacterial adhesion is to alter the surface properties of the medical devices and to make them less attractive for bacteria using surface coating techniques. For biomedical application the coatings should have both good biocompatible and mechanical properties. Diamond-like carbon (DLC) coatings have been attracting a great deal of interest for biomedical devices due to their excellent properties such as biocompatibility, haemocompatibility, low friction and chemical inertness (Maegawa et al. 2016). DLC is an excellent base coating to be alloyed with different elements. Incorporation of selected elements into DLC has emerged as an innovative approach to add unique functional properties to DLC coatings, thus opening up a range of new potential applications for biomedical devices (Dearnaley

and Arps 2005; Marciano et al. 2010; Maegawa et al. 2016; Swiatek et al. 2016). The incorporation of Si, F, N or Ag into DLC reduced bacterial adhesion (Liu et al. 2008; Marciano et al. 2010; Su et al. 2010; Cloutier et al. 2016). Moreover, Ti coatings improved the antibacterial property (Dunne et al. 2017). Recently, the incorporation of Ti into DLC has been shown to be an effective method to improve tribological properties (Bayóna et al. 2015). However no research has been reported on the influence of Ti-doped DLC coatings on bacterial adhesion.

In this study, a range of Ti-doped DLC coatings with different Ti contents were prepared on a stainless steel 316L substrate using a plasma-enhanced chemical vapour deposition technique. The influence of Ti contents in the coatings on bacterial adhesion was investigated with *Escherichia coli* and *Pseudomonas aeruginosa* which are commonly associated with infections of medical devices and implants (Rădulescu et al. 2016).

## Materials and methods

### Preparation of Ti-DLC coatings

In order to investigate the effect of Ti contents in DLC coatings on bacterial adhesion, a range of Ti-DLC

**CONTACT** Ying-Yu Zhao ✉ [zhaoyy2@cmu1.h.com](mailto:zhaoyy2@cmu1.h.com); Qi Zhao ✉ [Q.Zhao@dundee.ac.uk](mailto:Q.Zhao@dundee.ac.uk)

© 2018 The Author(s). Published by Informa UK Limited, trading as Taylor & Francis Group.  
This is an Open Access article distributed under the terms of the Creative Commons Attribution-NonCommercial-NoDerivatives License (<http://creativecommons.org/licenses/by-nc-nd/4.0/>), which permits non-commercial re-use, distribution, and reproduction in any medium, provided the original work is properly cited, and is not altered, transformed, or built upon in any way.

coatings with different Ti contents (1.3, 1.8 and 3.2 at.%, respectively) were produced on stainless steel 316L plates (25 mm × 26 mm × 1 mm) by a magnetron sputter system consisting of a Ti target combined with plasma-enhanced chemical vapour deposition (PECVD). The average coating thickness was controlled around 1 µm. The Ti contents in the Ti-DLC coatings were determined by Energy Dispersive X-ray analysis (EDX) (Philips XL 30, Eindhoven, Netherlands). A Dimension 3000 atomic force microscope (AFM Digital Instruments, Santa Barbara, CA, USA) was employed to characterise the surface topography of the coatings with a lateral resolution of 1–2 nm and a vertical resolution of 0.01 nm. Crosscut tests and abrasion resistance tests with water and sand at 5 bar under angle of 45° for 0.5 min were performed in order to evaluate the coating adhesion strength to the substrate.

### Contact angle and surface energy

The surface energy of a solid surface gives a direct measure of intermolecular or interfacial attractive forces. van Oss (1994) divided the total surface energy ( $\gamma_2^{TOT}$ ) of a solid into two components, Lifshitz-van der Waals (LW) apolar component ( $\gamma_2^{LW}$ ) and Lewis acid/base polar component ( $\gamma_2^{AB}$ ). The  $\gamma_2^{AB}$  is further divided into an electron donor ( $\gamma_2^-$ ) component and an electron acceptor ( $\gamma_2^+$ ) component. The contact angle on the coatings was measured using the sessile drop method with a Dataphysics OCA-20 contact angle analyser (DataPhysics Instruments GmbH, Filderstadt, Germany) with distilled water (W), diiodomethane (Di) and ethylene glycol (EG) as the test liquids, respectively. The surface energy components of the coatings were calculated using the van Oss approach (van Oss et al. 1994) using the contact angle data. The details of contact angle measurements and surface energy calculations were described in a previous study (Liu and Zhao 2011). In order to investigate bacterial adhesion mechanism, the surface energy of biofilms was also measured. To measure the contact angle of a biofilm, *E. coli* WT F1693 or *P. aeruginosa* ATCC 33347 were grown on a glass surface in bacterial suspension of  $10^6$  CFU ml<sup>-1</sup> for over 10 days with tryptone soya broth (TSB; Oxoid®, Cheshire, UK) as the growth medium. Using a microscope it was confirmed that a high density biofilm had formed on the glass slide at this time. The biofilm coated glass slide was removed from the bacterial suspension and was dried in air until it maintained a stable water contact angle. The contact angles of the living biofilm were then further measured by using the contact angle analyser with diiodomethane and ethylene glycol, respectively (Liu and Zhao 2011).

### Bacterial adhesion and removal

*E. coli* WT F1693 and *P. aeruginosa* ATCC 33347 were obtained from the Institute of Infection and Immunity,

Nottingham University, UK. Both *E. coli* and *P. aeruginosa* are microorganisms that frequently cause infections related to medical devices and implants. Each strain was sub-cultured and preserved in 15% glycerol in TSB (Oxoid®) as frozen stock at –80°C. For all adhesion tests, tryptone soya agar (TSA) plates were streaked out with a loop from the frozen stock and grown overnight at 37°C. A single colony was inoculated in 20 ml TSB and grown statically overnight at 37°C. From this culture 500 µl were further inoculated into 100 ml TSB in a conical flask and grown in a shaker-incubator at 37°C and 250 rpm. The culture was grown to mid-exponential phase (corresponding to an optical density absorbance at 600 nm of 0.5–0.7), since cells harvested in this growing phase have shown the best adhesion to a solid surface. Two ml of *E. coli* or *P. aeruginosa* were harvested by centrifugation at 2,500 rpm for 5 min at –4°C, washed once in sterile distilled water and resuspended in the 500 ml PBS (pre-warmed to 37°C) at a  $10^6$  CFU ml<sup>-1</sup> concentration.

For bacterial adhesion assays, six replicate samples of each coating were immersed in a glass tank containing 500 ml of the suspension of *E. coli* or *P. aeruginosa* ( $10^6$  CFU ml<sup>-1</sup>), incubated on a shaker (20 rpm) at 37°C for 1 h. Each sample was dipped twice vertically in a glass vessel containing 130 ml sterile distilled water with a custom-made automated dipper apparatus under a constant speed of 0.028 m s<sup>-1</sup> with corresponding shear stress of 0.014 N m<sup>-2</sup> in order to remove loosely attached bacteria. In order to observe the number of cells microscopically, the bacteria on the samples were stained using LIVE/DEAD BacLight Kit L 13152 for (Molecular Probes Europe BV, Leiden, The Netherlands) 15 min and then observed under a fluorescence microscope (OLYMPUS BX 41, Tokyo, Japan) and counted using Image Pro Plus software (Media Cybernetics, Rockville, MD, USA). Then the dead, live and total numbers of adhered bacteria on each sample were determined. For each sample, six measurements on the six different areas of the sample were made. The total measurements for each type of coating were 36.

To assess the adhesion strength of the attached bacteria, each sample was dipped a further 20 times vertically in the glass vessel containing 130 ml sterile distilled water at 37°C and a constant shear stress of 0.014 N m<sup>-2</sup>. Each test sample was stained and then observed under the fluorescence microscope to determine the remaining bacteria. Then the bacterial removal rate *R* (%) for the each sample was determined:

$$R = (A - L)/A \times 100\% \quad (1)$$

where *A* is total number of adhered bacteria before dipping process and *L* is the remaining bacteria after dipping process.

### Statistical analysis by *t*-test

The *t*-test is perhaps the most widely used method for comparing two independent groups of data. Each group is normally distributed around its respective mean value, and the two groups have the same variance. The sole difference between the groups is that their means may not be the same. The *t*-statistic is defined as:

$$t = \frac{\bar{x} - \bar{y}}{S \sqrt{1/n + 1/m}} \quad (2)$$

where  $\bar{x}$  is the sample mean of data in the first group  $x_i$ ,  $i = 1, 2, \dots, n$ ;  $\bar{y}$  is the sample mean of data in the second group,  $y_j$ ,  $j = 1, 2, \dots, m$ ; and  $S$  is the pooled sample standard deviation, estimating the standard deviation assumed identical in both groups:

$$S = \sqrt{\frac{(n-1)s_x^2 + (m-1)s_y^2}{n+m-2}} \quad (3)$$

The sample variances of both groups,  $s_x^2$  and  $s_y^2$  are used to estimate  $S$ .

$$s_x^2 = \frac{1}{n-1} \sum_{i=1}^n (x_i - \bar{x})^2 \quad (4)$$

$$s_y^2 = \frac{1}{m-1} \sum_{j=1}^m (y_j - \bar{y})^2 \quad (5)$$

The *t*-test was used as statistical analysis for evaluating the anti-microbial ability of coated and control surfaces. The conventional 0.05 level was considered to reflect statistical significance.

## Results

### Surface characterisation

The Ti contents in the three Ti-DLC coatings were determined by Energy Dispersive X-ray analysis, which were 1.3, 1.8 and 3.2 at.%, respectively. The surface topography and roughness of the coatings were measured with an AFM. The roughness for each surface was an average of six measurements at different position with surface area of  $5 \times 5 \mu\text{m}$ . Figure 1a, b, c and d show the AFM images of pure DLC coating, 1.3%Ti-DLC coating, 1.8%Ti-DLC coating and 3.2%Ti-DLC coating, respectively. Both pure DLC coating and the Ti-DLC coatings consisted essentially of hemispherical rounded hillock units, and their crystallites were distributed densely and homogeneously. Table 1 indicates that the surface roughness increases slightly with increasing Ti content in the DLC coatings. The results of the crosscut tests and abrasion resistance




Figure 1. AFM images of (a) DLC coating; (b) 1.3%Ti-DLC coating; (c) 1.8%Ti-DLC coating; (d) 3.2%Ti-DLC coating.



tests showed that the Ti-DLC coatings had very good adhesion to the substrates and very good resistance to abrasion.

### Contact angle and surface energy component

Table 2 shows the contact angle and surface energy components of coatings used in this investigation, including polished stainless steel, DLC coating, 1.3%Ti-DLC coating, 1.8%Ti-DLC coating and 3.2%Ti-DLC coating. The results showed that the  $\gamma_2^{LW}$  component of the Ti-DLC coatings almost kept constant, about 44 mJ m<sup>-2</sup>; while the  $\gamma_2^-$  component of the Ti-DLC coatings increased significantly with increasing Ti content in the coatings and the  $\gamma_2^+$  component of the coatings was almost zero. Table 2 also shows the contact angle and surface energy components of *E. coli* and *P. aeruginosa* biofilms used in this investigation.

### Bacterial adhesion and removal

Figure 2 shows the attachment of *E. coli* cells on stainless steel, DLC coating and Ti-DLC coatings. Due to exposure to air for 15 min, ~50% of the bacteria on the Ti-DLC coatings were dead. Both pure DLC coating and the Ti-DLC coatings performed much better than stainless steel against bacterial attachment. The 3.2% Ti-DLC performed best, which reduced total bacterial adhesion by 75% and 66%, respectively, compared with stainless steel and pure DLC coating. Figure 3 clearly shows that the number of adhered bacteria (live, dead and total bacteria) decreased with Ti content increasing with very high correlation coefficients ( $R^2=0.96-0.98$ ). The statistical analysis by *t*-test indicated that there was a significant statistical difference between the Ti-DLC coatings and stainless steel in reducing *E. coli* adhesion ( $p \leq 0.05$ ).

**Table 1.** Ti content and surface roughness Ra.

Coatings	Ra by AFM (nm)
DLC	4.7±0.4
1.3%Ti-DLC	7.9±0.9
1.8%Ti-DLC	10.6±1.2
3.2%Ti-DLC	17.3±1.4

During the dipping process, some bacteria were removed from the coatings. Figure 4 shows that the numbers of remaining bacteria (live, dead and total bacteria) decreased with Ti content increasing. Figure 5 shows that the removal percentage of *E. coli* cells from stainless steel, the DLC coating and the Ti-DLC coatings after dipping process increased with increasing Ti content in the DLC coating. The removal percentage of live, dead and total bacteria on stainless steel was only about 15%; while the bacterial removal percentage on the 3.2% Ti-DLC coatings increased to about 45%. The statistical analysis by *t*-test indicated that there was a significant statistical difference between the Ti-DLC coatings and stainless steel in the removal rates of *E. coli* cells ( $p \leq 0.05$ ).

### Influence of surface energy components


Figure 6 shows that the live, dead and total *E. coli* cells on the Ti-DLC coatings decreased with the  $\gamma^-$  surface energy of the Ti-DLC coatings increasing. Figure 7 demonstrates that the bacterial attachment had strong correlation with the  $\gamma_2^{LW}/\gamma_2^-$  ratio, which varied in a wide range between 4.52 and 10.44. The live, dead and total *E. coli* cells on the polished stainless steel, DLC coating and Ti-DLC coatings decreased with the  $\gamma_2^{LW}/\gamma_2^-$  ratio decreasing with the correlation coefficient  $R^2$  values 0.921, 0.903 and 0.914, respectively. Figure 8 shows that the bacterial removal percentage also had strong correlation with the  $\gamma_2^{LW}/\gamma_2^-$  ratio. The removal percentage of *E. coli* cells (live, dead and total *E. coli* cells) increased with the  $\gamma_2^{LW}/\gamma_2^-$  ratio decreasing with the correlation coefficient  $R^2$  values 0.972, 0.956 and 0.968, respectively.

### Influence of surface roughness


The effect of surface roughness on bacterial adhesion has also been investigated. These studies clearly demonstrate that substrate topography at the micro- and nano-scale had significant influence on bacterial attachment ( $p \leq 0.05$ ). However, no universal relationship between the surface roughness and the number of attached cells was observed (Hsu et al. 2013). Some studies showed that increased surface roughness at micro-scale leads to an increased bacterial adhesion, as roughening of the

**Table 2.** Contact angle and surface energy components of coating samples and bacteria.


Coatings or bacteria	Contact angle $\theta$ [°]			Surface energy components [mJ m <sup>-2</sup> ]				
	$\theta^W$	$\theta^{Di}$	$\theta^{EG}$	$\gamma^{LW}$	$\gamma^+$	$\gamma^-$	$\gamma^{TOT}$	$\gamma_2^{LW}/\gamma_2^-$
Polished steel	79.6±3.0	32.1±1.5	45.2±1.5	43.33	0.15	4.15	44.90	10.44
DLC	77.2±3.0	31.6±1.5	39.5±1.5	43.56	0.35	4.47	46.06	9.74
1.3%Ti-DLC	78.9±1.9	27.1±1.0	54.6±0.2	44.11	0.00	5.72	44.11	7.71
1.8%Ti-DLC	75.2±0.3	29.6±2.9	47.4±0.3	44.39	0.00	7.97	44.39	5.57
3.2%Ti-DLC	72.8±0.3	27.4±0.2	50.0±0.9	44.40	0.00	9.83	44.40	4.52
<i>E. coli</i>	17.2±1.3	78.0±2.1	18.6±0.6	18.52	3.05	70.01	47.75	—
<i>P. aeruginosa</i>	78.0±2.5	55.4±2.2	72.4±2.6	31.20	0.66	7.43	35.66	—



**Figure 2.** Attachment of *E. coli* cells on stainless steel, DLC coating and Ti-DLC coatings ( $N = 36$ , error bars are standard error).



**Figure 3.** Effect of Ti contents of Ti-DLC coatings on the attachment of *E. coli* cells ( $N = 36$ , error bars are standard error).



**Figure 4.** Effect of Ti content of Ti-DLC coatings on the remaining *E. coli* cells ( $N = 36$ , error bars are standard error).

surface increases the area available for bacterial adhesion (Gharechahi et al. 2012). Yoda et al. (2014) demonstrated that surface roughness ( $R_a$ ) below 30 nm can also promote bacterial adhesion. In this study, the  $R_a$  values of the Ti-DLC coatings were in the range of 4.7–17.3 nm (Table 1). Figure 9 shows that the number of attached bacteria decreased with increasing surface roughness

( $p \leq 0.05$ ). These findings are in accordance with the studies of Yoda et al. (2014).

### Influence of bacterial species

According the extended DLVO theory, the surface properties of bacteria, including surface charge and surface energy, have significant influence on bacterial adhesion (van Oss 1994; Liu and Zhao 2011). In order to validate the antibacterial properties of the Ti-DLC coatings, the assays of bacterial adhesion and removal were also performed with *P. aeruginosa*. Figure 10 shows the attachment of *P. aeruginosa* cells on stainless steel, DLC coating and Ti-DLC coatings; similar results to Figure 2 with *E. coli* were obtained. However it was observed that the total number of adhered *P. aeruginosa* cells was significant higher than that of adhered *E. coli* ( $p \leq 0.05$ ). Figure 11 shows the removal percentage of *P. aeruginosa* cells from stainless steel, the DLC coating and the Ti-DLC coatings. Again, similar results to Figure 5 with *E. coli* were obtained.

### Discussion

The influence of total surface energy on bacterial adhesion has been investigated intensively, with the frequent conclusion that bacterial adhesion is less on low energy surfaces and easier to clean because of weaker binding at the surface interface (Ahn et al. 2009; Liu and Zhao 2011; Brajkovic et al. 2014). Brajkovic et al. (2014) demonstrated that the surfaces with lower values of surface energy were less prone to biofilm adhesion and the number of adhered bacteria increased with increasing surface energy. However, there are also a number of contrary findings that suggest high energy surfaces have a smaller biofouling tendency and higher removal rate than low energy surfaces (Brink et al. 1993; Pasmore et al. 2001). Hahnel et al. (2009) reported that there was no relationship between total surface energy and bacterial adhesion. The above research indicated that there is no consistent conclusion on the effect of total surface energy of substrates on bacterial attachment and removal. Little research work has been reported on the effect of surface energy components of substrates on bacterial attachment and removal.

It was reported that the larger the electron donor component  $\gamma^-$  of a surface, the more negatively charged the surface (Chibowski et al. 1994; Liu and Zhao 2011). Therefore, bacterial adhesion should decrease with increasing electron donor  $\gamma^-$  values of the coatings if other parameters that affect bacterial adhesion are identical. The  $\gamma^-$  surface energy component of Ti-DLC coatings increased with Ti content increasing, meaning that the incorporation of Ti element increased the negative charge




Figure 5. Removal percentage of *E. coli* cells from stainless steel, pure DLC coating and Ti-DLC coatings after dipping process ( $N = 36$ , error bars are standard error).




Figure 8. Effect of  $\gamma_2^{LW}/\gamma_2^-$  ratio on the removal percentage of *E. coli* cells from stainless steel, DLC coating and Ti-DLC coatings after dipping process ( $N = 36$ , error bars are standard error).




Figure 6. Effect of  $\gamma^-$  surface energy of Ti-DLC coatings on the attachment of *E. coli* cells ( $N = 36$ , error bars are standard error).




Figure 9. Effect of surface roughness Ra of Ti-DLC coatings on the attachment of *E. coli* cells ( $N = 36$ , error bars are standard error).




Figure 7. Effect of  $\gamma_2^{LW}/\gamma_2^-$  ratio of stainless steel, DLC coating and Ti-DLC coatings on the attachment of *E. coli* cells ( $N = 36$ , error bars are standard error).





Figure 10. Attachment of *P. aeruginosa* cells on stainless steel, DLC coating and Ti-DLC coatings ( $N = 36$ , error bars are standard error).

of the Ti-DLC coatings. As *E. coli* and *P. aeruginosa* are negatively charged, this explains why the number of the adhered bacteria significantly decreased with increasing the electron donor  $\gamma^-$  values of the coatings and why the

Ti-DLC coatings performed much better than stainless steel or pure DLC coating against bacterial adhesion. The higher electron donor surface energy approach may lead to a way of controlling biofilm formation.

The comparison of the numbers of adhered bacteria shows that the numbers of *E. coli* was much lower than





**Figure 11.** Removal percentage of *P. aeruginosa* cells from Stainless steel, pure DLC coating and Ti-DLC coatings after dipping process ( $N = 36$ , error bars are standard error).

*P. aeruginosa*. The  $\gamma_2^-$  value of *E. coli* ( $70.01 \text{ mJ m}^{-2}$ ) is much higher than that of *P. aeruginosa* ( $7.43 \text{ mJ m}^{-2}$ ). This helps to explain why *E. coli* is relative more repellent to the coatings with high  $\gamma_2^-$  values.

## Conclusions

A range of Ti-DLC coatings with different Ti contents (1.3, 1.8 and 3.2 at.%) on stainless steel substrates were prepared using a plasma-enhanced chemical vapour deposition technique. The incorporation of Ti into the DLC coatings increased both the electron donor surface energy  $\gamma_2^-$  and surface roughness of the Ti-DLC coatings, which lead to the reduction of bacterial adhesion and the increase of bacterial removal rate. The 3.2% Ti-DLC coating performed best, which reduced bacterial attachment by 75% and increased bacterial detachment from 15% to 45%, compared with stainless steel. There were strong correlation between bacterial attachment (or removal) and the  $\gamma_2^{LW}/\gamma_2^-$  ratio. The  $\gamma_2^{LW}/\gamma_2^-$  ratio gives us a clear direction for the design of anti-biofouling and biofoul-release coatings through surface modification.

## Disclosure statement

No potential conflict of interest was reported by the authors.

## Funding

Financial support from the UK Engineering and Physical Sciences Research Council [grant number EP/P00301X/1].

## References

Ahn HB, Ahn SJ, Lee SJ, Kim TW, Nahm DS. 2009. Analysis of surface roughness and surface free energy characteristics

- of various orthodontic materials. *Am J Orthod Dentofacial Orthop.* 136:668–674. doi: [10.1016/j.ajodo.2007.11.032](https://doi.org/10.1016/j.ajodo.2007.11.032)
- Bayóna R, Igartua A, Gonzálezb JJ, Ruiz de Gopeguia U. 2015. Influence of the carbon content on the corrosion and tribocorrosion performance of Ti-DLC coatings for biomedical alloys. *Tribol Int.* 88:115–125. doi: [10.1016/j.triboint.2015.03.007](https://doi.org/10.1016/j.triboint.2015.03.007)
- Bociaga D, Komorowski P, Batory D, Szymanski W, Olejnik A, Jastrzebski K, Jakubowski W. 2015. Silver-doped nanocomposite carbon coatings (Ag-DLC) for biomedical applications – physicochemical and biological evaluation. *Appl Surf Sci.* 355:388–397. doi: [10.1016/j.apsusc.2015.07.117](https://doi.org/10.1016/j.apsusc.2015.07.117)
- Bradjkovic D, Antonijevic D, Milovanovic P, Kisic D, Zelic K, Djuric M, Rakocevic Z. 2014. Surface characterization of the cement for retention of implant supported dental prostheses: *in vitro* evaluation of cement roughness and surface free energy. *Appl Surf Sci.* 311:131–138. doi: [10.1016/j.apsusc.2014.05.027](https://doi.org/10.1016/j.apsusc.2014.05.027)
- Brink LES, Elbers SJG, Robbertsen T, Both P. 1993. The anti-fouling action of polymers preadsorbed on ultrafiltration and microfiltration membranes. *J Membr Sci.* 76:281–291. doi: [10.1016/0376-7388\(93\)85225-L](https://doi.org/10.1016/0376-7388(93)85225-L)
- Ceroni D, Grumetz C, Desvachez O, Pusateri S, Dunand P, Samara E. 2016. From prevention of pin-tract infection to treatment of osteomyelitis during paediatric external fixation. *J Child Orthop.* 10:605–612. doi: [10.1007/s11832-016-0787-8](https://doi.org/10.1007/s11832-016-0787-8)
- Chibowski E, Holysz L, Wójcik W. 1994. Changes in zeta potential and surface free energy of calcium carbonate due to exposure to radiofrequency electric field. *Colloids Surf A Physicochem Eng Asp.* 92:79–85. doi: [10.1016/0927-7757\(94\)02949-0](https://doi.org/10.1016/0927-7757(94)02949-0)
- Cloutier M, Turgeon S, Busby Y, Tatoulian M, Pireaux JJ, Mantovani D. 2016. Controlled distribution and clustering of silver in Ag-DLC nanocomposite coatings using a hybrid plasma approach. *ACS Appl Mater Interfaces.* 8:21020–21027. doi: [10.1021/acsami.6b06614](https://doi.org/10.1021/acsami.6b06614)
- Dearnaley G, Arps JH. 2005. Biomedical applications of diamond-like carbon (DLC) coatings: a review. *Surf Coat Technol.* 200:2518–2524. doi: [10.1016/j.surfcoat.2005.07.077](https://doi.org/10.1016/j.surfcoat.2005.07.077)
- Dunne CP, Keinänen-Toivola MM, Kahru A, Teunissen B, Olmez H, Gouveia I, Melo L, Murzyn K, Modic M, Ahonen M, et al. 2017. Anti-microbial coating innovations to prevent infectious diseases (AMICI): cost action ca15114. *Bioengineered.* 70:1–7. doi: [10.1080/21655979.2017.1323593](https://doi.org/10.1080/21655979.2017.1323593)
- Gharechahi M, Moosavi H, Forghani M. 2012. Effect of surface roughness and materials composition on biofilm formation. *J Biomater Nanobiotechnol.* 3:541–546. doi: [10.4236/jbnb.2012.324056](https://doi.org/10.4236/jbnb.2012.324056)
- Hahnel S, Rosentritt M, Handel G, Bürgers R. 2009. Surface characterization of dental ceramics and initial streptococcal adhesion *in vitro*. *Dent Mater.* 25:969–975. doi: [10.1016/j.dental.2009.02.003](https://doi.org/10.1016/j.dental.2009.02.003)
- Hsu LC, Fang J, Borca-Tasciuc DA, Worobo RW, Moraru CT. 2013. Effect of micro- and nanoscale topography on the adhesion of bacterial cells to solid surfaces. *Appl Environ Microbiol.* 79:2703–2712. doi: [10.1128/AEM.03436-12](https://doi.org/10.1128/AEM.03436-12)
- Liu C, Zhao Q. 2011. The CQ ratio of surface energy components influences adhesion and removal of fouling bacteria. *Biofouling.* 27:275–285. doi: [10.1080/08927014.2011.563842](https://doi.org/10.1080/08927014.2011.563842)

- Liu C, Zhao Q, Liu Y, Wang S, Abel E. 2008. Reduction of bacterial adhesion on modified DLC coatings. *Colloids Surf B Biointerfaces*. 61:182–187. doi: [10.1016/j.colsurfb.2007.08.008](https://doi.org/10.1016/j.colsurfb.2007.08.008)
- Maegawa S, Hasebe T, Yamato Y, Bitoa K, Nagashima S, Hayashi T, Mine T, Matsumoto T, Hotta A, Suzuki T. 2016. Time course analysis of antithrombogenic properties of fluorinated diamond-like carbon coating determined via accelerated aging tests: Quality control for medical device commercialization. *Diamond Relat Mater*. 70:33–38. doi: [10.1016/j.diamond.2016.09.021](https://doi.org/10.1016/j.diamond.2016.09.021)
- Marciano FR, Lima-Oliveira DA, Da-Silva NS, Corat EJ, Trava-Airoidi VJ. 2010. Antibacterial activity of fluorinated diamond-like carbon films produced by PECVD. *Surf Coat Technol*. 204:2986–2990. doi: [10.1016/j.surfcoat.2010.02.040](https://doi.org/10.1016/j.surfcoat.2010.02.040)
- Pasmore M, Todd P, Smith S, Baker D, Silverstein J, Coons D, Bowman CN. 2001. Effects of ultrafiltration membrane surface properties on *Pseudomonas aeruginosa* biofilm initiation for the purpose of reducing biofouling. *J Membr Sci*. 194:15–32. doi: [10.1016/S0376-7388\(01\)00468-9](https://doi.org/10.1016/S0376-7388(01)00468-9)
- Percival SL, Suleman L, Vuotto C, Donelli G. 2015. Healthcare-associated infections, medical devices and biofilms: risk, tolerance and control. *J Med Microbiol*. 64:323–334. doi: [10.1099/jmm.0.000032](https://doi.org/10.1099/jmm.0.000032)
- Rădulescu D, Grumezescu V, Andronescu E, Holban AM, Grumezescu AM, Socol G, Oprea AE, Rădulescu M, Surdu A, Trusca R, et al. 2016. Biocompatible cephalosporin-hydroxyapatite-poly (lactic-co-glycolic acid)-coatings fabricated by MAPLE technique for the prevention of bone implant associated infections. *Appl Surf Sci*. 374:387–396. doi: [10.1016/j.apsusc.2016.02.072](https://doi.org/10.1016/j.apsusc.2016.02.072)
- Su XJ, Zhao Q, Wang S, Bendavid A. 2010. Modification of diamond-like carbon coatings with fluorine to reduce biofouling adhesion. *Surf Coat Technol* 2010; 204:2454–2458. doi: [10.1016/j.surfcoat.2010.01.022](https://doi.org/10.1016/j.surfcoat.2010.01.022)
- Swiatek L, Olejnik A, Grabarczyk J, Jedrzejczak A, Sobczyk-Guzenda A, Kaminska M, Jakubowski W, Szymanski W, Bociaga D. 2016. Multi-doped diamond like-carbon coatings (DLC-Si/Ag) for biomedical applications fabricated using the modified chemical vapour deposition method. *Diam Relat Mater*. 67:54–62. doi: [10.1016/j.diamond.2016.03.005](https://doi.org/10.1016/j.diamond.2016.03.005)
- van Oss CJ. 1994. *Interfacial forces in aqueous media*. New York: Marcel Dekker.
- Yoda I, Koseki H, Tomita M, Shida T, Horiuchi H, Sakoda H, Osaki M. 2014. Effect of surface roughness of biomaterials on *Staphylococcus epidermidis* adhesion. *BMC Microbiol*. 14:97–240. doi: [10.1186/s12866-014-0234-2](https://doi.org/10.1186/s12866-014-0234-2)